MASTER

G€ND-INF- 017 Volume VIII January 1982



This is an informal report intended for use as a preliminary or working document



General Public Utilities • Electric Power Research Institute • U.S. Nuclear Regulatory Commission • U.S. Department of Energy

FIELD MEASUREMENTS AND INTERPRETATION OF

J. E. Jones J. T. Smith M. V. Mathis

Prepared for the U.S. Department of Energy Three Mile Island Operations Office Under DOE Contract No. DE-AC07-76ID01570

DISTRIBUTION OF THE STATE STATES

DISCLAIMER

This book was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately owned rights. References herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency the to review and opinions of authors expressed herein do not necessarily state or review those of the United States Government or any agency.

GEND-INF-017 Volume VIII

- DISCLAMER -

This because provided with end on the role of sponsors by an approval they unneed Steep Orientement Section this function. Steep Concentration and any steep steep from any of these methods and percentry, assess or important, an assessment any longer index to any of the approximation and percentry, assesses or important, and any steep terms to any steep readshifts, for the accuracy inducements, any performance of any index readshifts, and endowed the concentry and concentry products any other interviewed by steep terms table and the any steep terms to any specific concentry products or importantly to approximate the tradement and the approximation of the approximations of index readshifts, concentry or importantly table needs tradement and approximate the formation. They approximate the outportant and approximate or importantly table needs tradements and approximate the formation of the approximate the terms of index readshifts in any approximation the state of approximation of the approximate the approximate the state of the approximate terms of the state of

FIELD MEASUREMENTS AND INTERPRETATION OF TMI-2 INSTRUMENTATION: HP-R-212

GEND-INF--017 Vol. 8

J. T. Smith DE82 010679 M. V. Mathis

Technology for Energy Corporation

J. E. Jones

January 1982

Published on Behalf of the GEND Group by EG&G Idaho, Inc. Idaho Falls, Idaho 83415

> Prepared for the U.S. Department of Energy Three Mile Island Operations Office Under DOE Contract No. DE-AC07-76ID01570

> > DISTRIBUTION OF THIS ENGINEET IS DELINATED

TABLE OF CONTENTS

Sec	<u>ction</u>	Page
1.	INTRODUCTION	1-1
2.	INSTRUMENT LOCATION, CABLING, AND TERMINATIONS	2-1
3.	MEASUREMENT PROCEDURES	3-1
4.	MEASUREMENT S	4-1
5.	SUMMARY AND INTERPRETATION OF MEASUREMENTS	5-1
	5.1 Summary of Measurements 5.2 Interpretation of Measurements	5-1 5-6
6.	CONCLUSIONS	6-1

111

..

APPENDIX

LIST OF FIGURES

Figu	ire	Page
2-1	Area Monitor HP-R-212	2-2
2-2	Sketch of Instrumentation for HP-R-212	2-4
2-3	Functional Layout of Detector and Readout Module	2-5
2-4	Electrical Circuit of Detector Card	2-6
4-1	Typical Fluctuations Present on Checksource Line 1	4-2
4-2	Typical Fluctuations Present on Checksource Line 2	4-3
4-3	AC Variation on the 600V Power Supply	4-4
4-4	AC Variations on the Signal Output and a same	4-5
4-5	AC Variations on the 10V Power Supply	4-6
4-6	Typical Fluctuations Present Between Ground-AC Ground	4-7
4-7	High Frequency Spectrum of 10V Power Supply	4-8
4-8	High Frequency Spectrum of Signal Output	4-9
4-9	High Frequency Spectrum of 600V Power Supply	4-10
4-10	Low Frequency Spectra of 600V Power Supply	4-11
4-11	Low Frequency Spectra of Signal Output	4-12
4-12	Low Frequency Spectra of 10V Power Supply	4-13
4-13	TDR Traces from the Checksource Lines	4-16
4-14	TDR Traces from the 600V Power Supply Cable	4-17
4-15	5 TDR Traces from the Signal-In Cable	4-18
4-16	TDR Traces from the 10V Power Supply Lines	4-19

tv

LIST OF TABLES

Iddi	<u>e</u>	Page
2-1	Termination Points for HP-R-212 Measurements	2-3
4-1	Capacitance, Impedance, and Resistance Measurements	4-15
5-1	Characteristics of Major Signal Lines	5-2
5-2	Comparison of Measured Detector Resistance	5-4
5-3	Summary of TDR Measurements	5-5
×		

i di s

1. INTRODUCTION

During and following the TMI-2 accident, a number of instruments failed or were suspected of providing erroneous readings. Because of this problem, industry concerns were focused upon the behavior of instrumentation under adverse conditions. To better understand failure mechanisms, the Technical Integration Office (TIO) contracted Technology for Energy Corporation (TEC) to perform field measurements on a set of selected TMI-2 instruments to determine in-situ operating characteristics. For some instruments, these measurements were to be performed prior to removal (and replacement with new instruments) in order to have a cross reference with post-removal observations. For other instruments, an indication of the condition of the instrument (i.e., fully operational or failed) was desired.

This report provides the information gathered by TEC on the area radiation monitor HP-R-212. This detector was located at 305 feet elevation inside containment. This instrument consisted of a Victoreen Model 857-2 detector assembly connected to a Victoreen Model 856-2 panel alarm and approximately 1200 feet of interconnecting cable. This instrument was believed to have failed due to a constant 45 mR/hr radiation level indication and due to a lack of response to the manually activated checksource in the detector. As a result of this failure, the detector was a candidate for early replacement to provide long-term radiation monitoring capability inside containment.

2. INSTRUMENT LOCATION, CABLING, AND TERMINATIONS

A review of appropriate drawings from Victoreen and Burns & Roe (itemized in the Appendix in the measurement procedure, page A-5) resulted in the composite electrical diagram shown in Figure 2-1. From this information, Table 2-1 gives a list of the appropriate termination points for performing measurements in the Control Room in Cabinet 12. Also noted in Figure 2-1 are the cable lengths pulled during instrument installation (before final trimming) between each termination and/or junction point.

The detector assembly is a Victoreen Model 857-2 which is shown in Figure 2-2 along with required interfacing connections to the readout module. Figure 2-3 shows the functional layout of the detector and associated readout module. This assembly is a "medium range" device with a range of 0.1 to 10^4 mR/hr. An electrical diagram of the detector circuit is shown in Figure 2-4. As shown in Figure 2-1, the circuit is somewhat complicated by the presence of a remote alarm/meter and a second remote alarm which are used as local indicators of the radiation levels.

Since measurements were being made in the control room, there was no way to remove the effect of the remote alarm/meter (attached to the signal line) from the observed instrument response. However, since the remote alarm/meter was located outside containment, it did not experience the severe operating environments and thus was not considered to present any measurement problems. (The remote alarm was located inside containment,





2	-	3	

Table 2-1

TERMINATION POINTS FOR HP-R-212 MEASUREMENTS

Si gna 1	Cabinet 12 Identification*
+10V Power Supply	TB110-8
+600V High Voltage	TB110-5
Signal In	TB110-6
Ground	TB110-10
C S**	TB110-1
CS**	TB110-2

*From cable IT1870I

**CS = Checksource coil positive and return contacts (exact identification
 not necessary).



Figure 2-2. Sketch of Instrumentation for HP-R-212.

2-4

国住于1月1



Figure 2-3 Functional Layout of Detector and Readout Module.





but was isolated from the signal line by the remote alarm/meter circuitry.) Similarly, the Model 356-2 Readout Module, located in the control room, was not specifically considered to be a source of instrumentation problems except in its function of supplying power to the detector assembly.

3. PREPARATION OF MEASUREMENT PROCEDURES

As a result of generating the composite electrical diagram and from a review of the Victoreen Area Monitor Operation Manual, the major types of measurements to be performed were identified as:

- 1. Determine as-found condition of Readout Module and Remote Meter and record signal output
- Perform passive measurements (i.e., passively monitor signals) on each electrical connection consisting of time domain waveforms, very-high frequency spectrum analysis (i.e., MHz region), and frequency spectra below 100 kHz
- 3. Perform resistance, capacitance, impedance, and Time Domain Reflectometry (TDR) active measurements (i.e., actively introducing a test signal).

These measurements were designed to verify the operation of the Readout Module (especially the power supplies), but the focus of the measurement was on the detector assembly, cabling, and terminations/connections to the assembly. The Appendix contains the detailed procedure which was followed during the measurement program, and a summary of measurements is presented in the next section.

3-1

4. MEASUREMENTS

Since the output of HP-R-212 was designed to cover the range of 0 to +10 volts, the signal could be directly measured without amplification. Before performing measurements, the readout of HP-R-212 indicated 45 mR/hr for the gamma dose inside containment. Activation of the checksource had no effect on the output reading. The Signal In was then recorded for approximately 10 minutes on an FM recorder and various outputs measured with a DVM. These measurements yielded the following results:

10 V power Supply @ 10.1 V Signal Out @ 4.3 VDC 600 V Power Supply @ 469 VDC @ 599 VDC (no load)

Checksource @ 13.8 ma.

The next measurements consisted of photographing the output waveforms of the checksource, Signal In, and power supplies from a storage oscilloscope. Figures 4-1 to 4-6 show the results of these time trace measurements. Along with the time traces, both high and low frequency spectra (frequency domain) were taken of the Signal In and power supplies. Figures 4-7 to 4-9 show the measured spectra over high frequency bandwidths (>1 MHz), while Figures 4-10 to 4-12 show spectra over bandwidths below 100 kHz.

Following the frequency spectra measurements, electrical calibration was performed on the HP-R-212 readout module by a TMI technician. No significant adjustments were noted during this calibration. (See calibration



Photo # 102-1 Time = 5msec Gain = 0.1V/div

Figure 4-1. Typical Fluctuations Present on Checksource Line 1.

• 7



Photo #102-2 Time - 2msec Gain - 0.1v/Div

Figure 4-2. Typical Fluctuations Present on Checksource Line 2.



Photo #102-3 Time - 50µsec Jain - 10my/div



Photo #102-4 Time - 5msec Sain - 10mV/div



Photo #102-5 Time - 0.1msec Sain - 20mV/div

Figure 4-3. AC Variations on the 600V Power Supply.

4-5



Photo #102-6 Time - 5µsec Gain - 2V/div



Photo #102-7 Time - 10µsec Gain - 2V/div

Figure 4-4. AC Variations on the Signal Output.



4-6

Photo #102-8 Time - 10µsec Jain - 50mV/div

,



Photo #102-9 Time - 2maec Gain - 50mV/div

Figure 4-5. AC Variations on the 10V Power Supply.









Figure 4-8. High Frequency Spectrum of Signal Output.







Frequency (kHz)



100 kHz F-MAX

+10db Reference

20 kHz harmonics illuminated

(16 kHz also piesent)

4 13



Photo	#102-16	
l kHz	F-MAX	
+10db	Reference	
60 Hz	harmonics	illuminated

Figure 4-10. Low Frequency Spectra of 600V Power Supply:



Photo #102-17 100 kHz F-MAX +10db Reference 40 kHz harmonics illuminated



1.1

Photo #102-18

1 kHz F-MAX

+10db Reference

1⁴ .

60 Hz harmonics illuminated

Figure 4-11. Low Frequency Sprectra of Signal Output.







Photo #102-19 100 kHz F-MAX +10db Reference

20 kHz harmonics illuminated



data sheet in the Appendix on page A-20.) After electrical calibration, power was removed from HP-R-212. The test fixture was removed and all signal lines from cable IT1870I to cabinet 12 were disconnected.

A series of active measurements (i.e., actively introducing a test signal into the circuit) was then performed. Table 4-1 shows the results of capacitance, impedance, and DC resistance measurements on some of the field cable lines (see Appendix pages A-14 and A-16 for a complete set). A set of TDR measurements was taken on the signal lines to determine possible cable defects. These TDR traces are shown in Figures 4-13 to 4-16.

Table 4-1

	C	Capacitance (nF)		Impedance (ohms)		Resistance	
Signal	100 Hz	1 kHz	.100 kHz	100 Hz	1 kHz	100 kHz	(ohms)*
Checksource (+) Checksource (-)	-9.84 μF	-5₊44 µF	838	69.1	75.4	4.23	38 (40)
+600 V Ground	-400	22	-3.8 μF	VAR**	VAR	2.08	>2 M
Signal In Shield	VAR	30	-4 μF	4.9k	4.9k	1.88	9.94k (7.4k)
+10 V Ground	6 μF	4.6 μF	186	250	51	30	15.2k (7.7k)
Ground (field) Ground (cabinet)	VAR	-76 μF	43	1. 1 .	4.4	39	
Signal In 10 V				. 			8.7k (7.1k)

4-15

CAPACITANCE, IMPEDANCE, AND RESISTANCE MEASUREMENTS

*Values in parentheses are reverse polarity values.

**Indicates variable response.

4-16

STRIP CHART 102-1

TB110: 1 to 2

Signal - Check Source



' `.







Setting - 500mp/div Range - 52.6 ft/div Sensitivity - 0.5 15 hz filter 2nd plot begins @ 800 ft. Cable dielectric - other





STRIP CHART 102-3 TB110: 6 to 7 Signal - Signal In



Setting - 500mp/div Range - 52.6 ft/div Sensitivity - 0.5 15 hz filter 2nd plot begins @ 800 ft. Cable dielectric - poly

en an Artista (1993) An Artista (1994) An Artista (1994) An Artista (1994)



Figure 4-15. TDR Traces from the Signal-In Cable.

STRIP CHART 102-4 TB110: 8 to 10 Signal - +10V



Aller Chill

オイン 手 かくびっき 正知

Setting - 500mp/div Range - 52.6 ft/div Sensitivity - 0.5 15 hz filter 2nd plot begins @ 800 ft. Cable dielectric - poly



Figure 4-16. TDR Traces from the 10V Power Supply Lines.

5. SUMMARY AND INTERPRETATION OF MEASUREMENTS

This section presents a summary of the interpretation of the measurements taken on HP-R-212. This interpretation is intended to indicate the condition of the device based on observed data.

5.1 SUMMARY OF MEASUREMENTS

The 10 V power supply measurements indicated a 10.1 VDC output value, which is within the expected range. The 600 V power supply was measured at 469 VDC when connected to the detector assembly, but 599 VDC when the connecting cable to the assembly was removed. This indicates an excessive load on the high voltage due to some problem in the detector or cable, or a defective power supply. The Signal In measurement produced a 4.3 VDC reading, which is lower than the expected value of 5.0 VDC for a 0-10 V pulsing output. Measurements on the checksource produced a 13.8 ma current load, which indicates the electrical path through the checksource coil is intact.

The time traces and frequency spectra were used to summarize the major characteristics of the measured waveforms. Results of this summary are presented in Table 5-1. Both the power supplies exhibit normal characteristics when compared to other TMI-2 measurements. The only indication of a possible problem is from the relatively large 100 mV P-P 120 hertz ripple in the 10 V supply; however, this is not believed to be large enough to cause failure of the instrument.

Table 5-1

Signa l	Frequency	Amplitude
Signal Out	0 (DC) 60 Hz and Harmonics 40 kHz	4.3 V 5.6 mV RMS 1.1 V RMS
	Total Spectrum	6 V P-P (1.3 - 7.3 V)
600 V Supply	60 Hz 16 kHz and Harmonics 20 kHz and Harmonics	12.7 mV RMS <1 mV RMS 4 mV RMS
· ·	Total Spectrum	60 mV P-P
10 V Supply	120 Hz 20 kHz and Harmonics	100 mV P-P 32 mV RMS
	Total Spectrum	200 m¥ P-P

CHARACTERISTICS OF MAJOR SIGNAL LINES

The Signal In waveform would normally be 0-10 V pulses with a frequency proportional to the radiation present. The oscilloscope photograph of the Signal Out line (Figure 4-4) shows that the output is a periodic waveform of approximately 6 volt peak-peak variation (and a 4.3 volt offset). Frequency spectra show the frequency of the signal to be 40 kHz and nearly sinusoidal. Resistance data (Table 4-1) similarly indicate a problem among the Signal In, +10V, and Shield lines. Table 5-2 gives a comparison of resistance data from a new detector assembly to that from HP-R-212. All resistances are higher than for the new assembly, but only the Signal In-Shield (+) and the Shield-10 V (-) show a significant increase.

The capacitance and impedance data given in Table 4-1 is difficult to quantitatively interpret due to active components in the circuitry, but qualitative results are possible. Very low effective capacitance values would be expected from most signal lines except for the +10 V to ground, which has a 100 μ F capacitor present. The checksource lines indicate the presence of the coil inductance (negative capacitance) which is also expected. Impedance data is reasonable and exhibits major trends expected from the circuitry such as reducing values at higher frequency for the Signal In.

The results of TDR measurements performed on the cable (shown in Figures 4-13 to 4-16 are summarized in Table 5-3. Note that the lengths identified in the table are only approximate, since no calibration of the cable resistance and material composition was performed on the TDR instrument. Some junction points were not identified by these

5-4

Table 5-2

COMPARISON OF MEASURED DETECTOR RESISTANCE

Measurement	New Detector*		HP-R	HP-R-212	
Point	Polarity +	Polarity -	Polarity +	Polarity +	
Signal In Shield	8.78	7.25	9.94	7.40	
Signal In +10 V	8.32	6.70	8.70	7.10	
Shield +10 V	7.40	11.80	7.70	15.20	

*Serial numbers 111 and 1405 composite data.

- Notes: (a) All values are in ohms $\times 10^3$ unless otherwise indicated.
 - (b) First signal to positive terminal, second signal to negative terminal is considered Polarity +.

۰ ·

(c) All measurements made with a Keithley 177 DVM or 20×10^3 ohm scale.

5-5 Table 5-3

......

SUMMARY OF TDR MEASUREMENTS

Signal Lines	Distance (ft)*	Description**	Probable Cause
Checksource (+) Checksource (-)	505 996 1285	Point R Increase Point R Increase Continuous R Increase	Remote Meter Penetration R507 HP-R-212 Detector
600 V Supply Ground	515 1246	Point R Increase Continuous R Increase	Remote Meter HP-R-212 Detector
Signal Shield	505 1320	Point R Increase Continuous R Increase	Remote Meter HP-R-212 Detector
10 V Supply Ground	289 474 975 1274	Point R Increase Point R Increase Point R Increase Continuous R Decrease	(?) Remote Meter Penetration R507 HP-R-212 Detector

Note: Distances are not calibrated due to lack of prior information on the cable type which prevented calibration tests.

*TDR to terminal block test cable (15 ft) not included in distance. **R is the abbreviation for resistance. measurements, but this is not unusual due to the cable lengths involved and the small resistance changes that would occur at terminal block junctions. The only unusual inflection occurred on the 10 V power supply line which indicated an interface resistance at approximately 289 feet from the control room. However, measurements did not indicate a resistance change on the 10 V line, and this was not considered to be a problem.

5.2 INTERPRETATION OF MEASUREMENTS

Based upon the observation of a periodic 40 kHz output on the Signal In line from the detector and the excessive loading on the 600 V power supply, it appears that the Geiger tube has failed. If the tube failed in an ionized condition (i.e., depletion of the quench gas), there would be an excessive current load on the 600 V power supply which could result in the observed drop in the supply voltage. Similarly, the detector assembly contains an "anti-jam" circuit which is designed to produce a periodic output upon saturation of the Geiger tube. This was designed to prevent loss of signal in the event of over-range radiation levels, but would also be triggered by a "shorted" tube.

In addition, the output of the detector should consist of 0-10 V pulses even if the "anti-jam" circuit is active. The observed output was nearly sinusoidal, ranging from approximately 1.3 V to 7.3 V, and resistance measurements indicated an increase in the expected values. A review of the detector circuit shows that this behavior could be explained by a failure of transistor Q7. The measurements indicate that

5-6

. Salation

6. CONCLUSIONS

Based on the measurements, data reduction, and circuit analysis of HP-R-212, there is an indication of failure of the instrument. The observed output signals and resistance measurements suggest that the Geiger tube is in a continuously ionized state and that the output is being generated by the "anti-jam" circuit. Also, there is an indication that one of the output driver transistors, Q7, has failed. The result of these failures is a nearly sinusoidal 40 kHz output that spans approximately 6 volts peak-peak and is erroneously interpreted as a 45 mR/hr detector response by the readout module.



Control Room Panal 12, Solder. Location:

Limits and Precautions: a) Personnel

b) Equipment

٩

c) Environment

d) Nuclear

Post Maintenance Testing required and Acceptance Criteria.

ORIGINATOR -- SUPERVISOR -- SUPERVISOR OF MAINTENANCE -- MAINTENANCE FOREMAN --JOB PERFORMER -- MAINTENANCE FOREMAN -- SUPERVISOR OF MAINTENANCE

Page A-2 WORK REQUEST PROCEDURE **TMI Nuclear Station** Maintenance Procedure Format and Approval Unit No.____Z This form outlines the format and acts as a cover sheet for a maintenance procedure. Due to the limited size of the form, additional pages may be attached as required. Work Request procedure AP 1016 Section 6 should be used as a guide in preparing the maintenance procedure. Procedure Title & No.: 1. Cable & Detector Performanse Check for HR R. 212 Purpose: To determine if imjesser reading als to cable a detector problem. 2. Description of system or component to be worked on. 3. HP-L-212 4. References: Victoren Manual Special Tools, and Materials required. 5. See attached 6. Detailed Procedure (attach additional pages as required) See attached Supervisor of Maintenance recommends approval <u>Elected</u> Date <u>9/16/80</u> ROVAL Engineering Levres Jet Brumme Pate 9/16/80 * PORC RECOMMENDS APPROVAL Unit No. 1 Chairman_____ Date_____ Unit No. 2 Chairman_____ Date_____ * UNIT SUPERINTENDENT APPROVAL _____ Date_____ Unit No. 2 _____ Date ____ Unit No. 1 * Standing Procedure Supervisor of QC Date *Note: These approvals required only on Nuclear Safety Related/Radiation work permit jobs. TMI-94 2.78

			r a ^r
·			÷
· .		Page A-3	
		TITLEIN-SITU MEASUREMENTS OF CABLES AND	NO.
•		HP-R-212	REV. 0
	Technology for Energy Corporation	APPROVED If. U. Martin	DATE
ſ	PROCEDURE	M.V. Mathis, Director, Tech. Serv. Div.	9-12-80
	mation in preparat Radiation Monitor tests specified in the in-containment cabling, and reado Time Domain Reflec (frequency domain) oscilloscope obser under test (UUT).	tion for possible removal and replacement of / HP-R-212 from the reactor building TMI Unit 2 in this procedure are designed to assess the co instrument module (gamma detector), associate but devices. This assessment will require the ctometry (TDR), Impedance (Z), Spectral Analys), special calibration measurements, and gener rvations (with recording) of waveforms from/to	Area 2. The ondition of ted tuse of tis tal the unit
	PROCEDURE (ADMINISTRATIVE):		· · ·
	A. Limitations and Pre	ecautions	,
	 <u>Nuclear Safety</u> dant ARM system the engineered relevance. 	 Area radiation monitor HP-R-212 is part of n at elevation 305'. The unit is not consider reactor safeguards system thus has no nuclear 	a redun- red part of safety
	 Environmental S out-of and rest environment. 	Safety. Area radiation monitor HP-R-212 can b tored to service without producing a hazard to	be taken b the
	 Personnel Safet personnel safet forming instrum 	<u>ty</u> . The test described herein produces no add ty hazards other than normally associated with ment calibrations and tests.	litional 1 per-
	4. <u>Equipment Prote</u> herein, care wi follows:	ection. In the performance of each test descr 11 be taken to insure adequate equipment prot	ibed ection as
	 a. In all case shall be ma Rfcowwection b. All passive observation shall be performed 	es actual test hookups to the Unit-2 instrumer ade and verified by Instrumentation Personnel. NJ 32 VELIFIED BY IIC PEES- e measurements (Spectral Analysis and Oscillos ns) of waveforms and signals from powered inst erformed using high input impedance probes or	tation AND cope ruments jnputs
	c. In all Time will be ren signals pre	e Domain Reflectometry and Impedance measureme moved from the unit under test and low level t escribed in Table 4-1 shall be utilized to per	ents, power cest form cable

Page A-4

IN-SITU MEASUREMENTS OF CABLES AND SIGNALS TITLE FROM AREA RADIATION MONITOR HP-R-212

REV, 0

integretary measurements on the appropriate instrumentation cables by inserting test signals on appropriate conductors of Cable IT1870I (terminations shall be removed and replaced on TB110 of Cabinet 12). Should these tests reveal cable integretary problems further verification measurements will be made at TB1 of the appropriate Remote Alarm/Meter (Victoreen Model 858-3) located in the anteroom.

Table 4-1 Active Measurements

Active Signal Parameter	Time Domain Reflectometry	Impedance
Voltage Frequency Current	225 mY nominal (into 50 ohm base) < 10mA	<pre></pre>
Other	225mV, 110 picosecond pulses	

d. In the calibration verification measurements section, baseline data on the as-found condition will be recorded prior to the performance of any adjustments or electronic calibrations.

- B. Prerequisites
 - 1. The Shift Supervisor/Shift Foreman shall be notified for concurrance prior to the performance of those measurements.
 - Instrumentation personnel shall be assigned to assist in the performance of these measurements.
 - 3. All measurements and test instrumentation shall be in current calibration (traceable to NBS).

PAGE 2 of 15

	IN-SITU MEASUREMENTS OF CABLES AND SIGNALS	NO. TP-102
	TITLE FROM AREA RADIATION MONITOR HP-R-212	REV.
4.	The Shift Supervisor/Shift Foreman shall be notified prior and upon completion of the measurements.	to starting
C. Pro	ocedure for Performing Measurements	
Ret	ferences :	
1.	Victoreen Dwg. No. 904550, Wiring Diagram Area Monitors Cha HP-R-211 & HP-R-212 (Sheet 5 of 11).	nnels
2.	Instruction Manual for G-M Area Monitoring Systems, Model 8 Victoreen Part No. 855-10-1.	55 Series
3.	Burns & Roe Dwg. 3347, Sh. 6L and Sh. 6J.	
4.	Burns & Roe Dwg. 3043, Sh. 16D.	¢
5.	Burns & Roe Dwg. 3045, Sh. 34C.	
6.	Burns & Roe Dwg. 3045, Sh. 34A.	
7.	Instruction Manual, Tektronix model 1502 Time Domain Reflec	tometer.
8.	Instruction Manual, Hewlett Packard Model 4274 Multifrequen Meter.	cy LCR
9.	Instruction Manual, Hewlett Packard Spectrum Analyzer (Mode 85538, 8552B Modules).	1 141T,
10.	Instruction Manual, Nicolet Model 444A-26 Spectrum Analyzer	•
. 11.	Instruction Manual, Tektronix Model 335 Oscilloscope.	x
12.	Instruction Manual, Lockheed Store-4 Recorder.	
13.	Instruction Manual, Tektronix SC502 Oscilloscope.	-
14.	TEC Composite Electrical Connection Diagram, HP-R-212 (see	attached).
Vic (Re of	ctoreen Instrument Company Dwg. 904550 (Ref. 1) and B&R Drawi ef. 3) show the appropriate termination points for passive me signals from HP-R-211 as follows:	ngs 3024 asurements
		:

ういしょう 人言が音 清重さ

	IN-SIT	U MEASUREMENTS	OF CABLES AND SIGNALS	NO. TP-102
	TITLE FROM A	REA RADIATION	MONITOR HP-R-212	REV.
	Signal	Cable IT2994I	Cabinet 12	
	+10V		TB110-8	
	600V		TB110-5	
	SIG		TB110-6	
	GND		TB110-10	÷
×.	cs		TB110-1	
			TP110.2	

<u>STEPS</u>

1. Notify Shift Supervisor/Shift Foreman of start of test on HP-R-212.

2. Verify power is applied to HP-R-212.

17/50 Signature/Date

3. Record present signals and readings and indications on 856-2 Readout Module (Local & Remote). Record Signal-in at TB110-6/7 and record output for 30 minutes on FM Tape Recorder. Remove recorder when finished.

Page A-7 NO. TP-102 IN-SITU MEASUREMENTS OF CABLES AND SIGNALS TEL TITLE FROM AREA RADIATION MONITOR HP-R-212 REV. n <u>Local</u> Meter/Indicator/Switch Rmte N/A **A**
(mR/hr Meter Reading Off-Operate-Alarm Function Switch OPERATE N/A Off Fail Safe Indicator 0n N/A Off High Alarm-Reset Indicator 0n N/A TEC, ignature/Date **9/11/5**0 4. Using a Keithley Model 177 DMM (or equivalent) and an electrostatic voltmeter $(Z_j \ge 10^{12} \text{ OHMS}, \text{ Range } 0-2000 \text{ V}, \text{ Precision } = \pm 1\%)$ measure the DC voltage or current at the following test points. NOTE: For signal d. it will be necessary to depress Fall-Safe Check Source push button during the measurement. Note: GPU Flake Differential Voltmeter Je Jun PAGE 5 of 15



	TITL	IN-SITU E FROM ARE	MEASUREMENTS OF A RADIATION MONI	CABLES AND SIGNA Tor HP-R-212	LS TP-102
5. Using form	a Tektronix at the follow	Model SC502 ing test poi	(or equivalent) nts:	oscilloscope obs	erve the wave-
SIGNAL	CABINET 12	PARAMETER	• 		
a.	TB110-1 TB110-10	CS	Photo <u>102-1</u> Time Base <u>5ms</u> ; Vert Gain <u>17</u>	Photo Time Base Vert Gain	Photo Time Base Vert Gain
b.,	TB110-2 TB110-10	ĊS	Photo <u>102-1</u> Time Base <u>2m5</u> Vert Gain <u>IV</u>	Photo Time Base Vert Gain	Photo Time Base Vert Gain
. *c.	TB110-5 TB110-10	+600V	Photo <u>/02-3</u> Time Base <u>50µ5</u> Vert Gain <u>/Omv</u>	Photo <u>102-4</u> Time Base <u>5 mS</u> Vert Gain <u>10 mY</u>	Photo <u>192-5</u> Time Base <u>Lass</u> Vert Gain
d.	TB110-6 TB110-7	SIG	Photo <u>/02-6</u> Time Base <u>5:5</u> Vert Gain <u>27</u>	Photo <u>102-7</u> Time Base <u>1015</u> Vert Gain <u>27</u>	Photo Time Base Vert Gain
	TB110-8 TB110-10	+10V	Photo <u>102-8</u> Time Base <u>1045</u> Vert Gain <u>50m</u> ¥	Photo <u>102-9</u> Time Base <u>2m5</u> Vert Gain <u>50m</u> V	Photo Time Base Vert Gain
f.	TB110-10 TB501-30	GND ACGND	Photo <u>102-10</u> Time Base <u>2005</u> Vert Gain <u>27</u>	Photo <u>/02-//</u> Time Base <u>5m5</u> Vert Gain <u>27</u>	Photo Time Base Vert Gain

*Decouple DC Voltage.

Sync the oscilloscope and photograph the waveform using up to three time base and vertical gain settings. (The necessity of 3 photographs will be determined by visual analysis by the field engineer.) Mark the back of the photographs with the instrument tag number and parameter measured.

(TEC) 9/17/80 Signature/Date

PAGE _7 of 15

	דודו ב	IN-SITU MEASUR	EMENTS OF CAR	LES AND SIG	INALS	TP-102
a dagan dagan	1 1 4 4 44444 	FRUM AREA RADI			17	EV. 0
		. [.] .				
6. Using equiva	a Hewlett-Packar lent) perform ar	rd Spectrum Ana a analysis of t	lyzer (Models he following	141T, 8553 signals for	B, and 8 spectra	552, or 1 content:
	SIGNAL	CABINET 12	PARAMETER	<u>PHOTO #</u>		
	a.	TB110-8 TB110-10	+10V GND	102-12	Bis: 3ch Atten: Oc 10 dB	Hz, SOCHHE/DA
	b.	TB110-6 TB110-7	SIG IN GND	102-13		/1
	*c.	TB110-5 TB110-10	+600V GND	102-14		11
	*Decouple (50VDC)	DC voltage max	input to Spe	ectrum Analy	zer	
Before resolu width	*Decouple (50VDC) photographing e tion. Record ce and sweep speed	DC voltage max each scope disp ritical analyze on rear of pho	input to Spe lay adjust ar r parameters tograph as we	ectrum Analy alyzer for e.g., RF ba il as param	best spe indwidth, meter ana	ctral RF band- lyzed.
Before resolu width	*Decouple (50VDC) photographing e tion. Record cr and sweep speed <u>SPECTRUM ID</u>	DC voltage max each scope disp ritical analyze on rear of pho ENT FREQUENC	input to Spe lay adjust ar r parameters tograph as we <u>Y AMPLITUE</u>	ectrum Analy alyzer for e.g., RF ba 11 as param DE <u>REMARK</u>	best spe indwidth, weter ana	ctral RF band- lyzed.
Before resolu width	*Decouple (50VDC) photographing e tion. Record cr and sweep speed <u>SPECTRUM ID</u>	DC voltage max each scope disp ritical analyze on rear of pho ENT FREQUENC	input to Spe lay adjust ar r parameters tograph as we <u>Y</u> <u>AMPLITUE</u>	ectrum Analy alyzer for e.g., RF ba 11 as param DE <u>REMARK</u>	best spe indwidth, weter ana	ctral RF band- lyzed.
Before resolu width	*Decouple (50VDC) photographing e tion. Record cr and sweep speed <u>SPECTRUM ID</u>	DC voltage max each scope disp ritical analyze on rear of pho ENT FREQUENC	input to Spe lay adjust ar r parameters tograph as we <u>Y</u> <u>AMPLITUE</u>	ectrum Analy alyzer for e.g., RF ba all as param DE <u>REMARK</u>	best spe indwidth, meter ana	ctral RF band- lyzed.
Before resolu width	*Decouple (50VDC) photographing e tion. Record cr and sweep speed <u>SPECTRUM ID</u>	DC voltage max each scope disp ritical analyze on rear of pho ENT FREQUENC	input to Spe lay adjust ar r parameters tograph as we <u>Y</u> <u>AMPLITUE</u>	ectrum Analy alyzer for e.g., RF ba all as param DE <u>REMARK</u>	best spe indwidth, meter ana	ctral RF band- lyzed.
Before resolu width	*Decouple (50VDC) photographing e tion. Record cr and sweep speed <u>SPECTRUM ID</u>	DC voltage max each scope disp ritical analyze on rear of pho ENT FREQUENC	input to Spe lay adjust ar r parameters tograph as we	ectrum Analy alyzer for e.g., RF ba ell as param DE <u>REMARK</u>	best spe indwidth, heter and SS	ctral RF band- lyzed. TC) 9/17/50
Before resolu width	*Decouple (50VDC) photographing e tion. Record cr and sweep speed <u>SPECTRUM ID</u>	DC voltage max each scope disp ritical analyze on rear of pho ENT FREQUENC	input to Spe lay adjust ar r parameters tograph as we <u>Y AMPLITUE</u>	Analy Analyzer for e.g., RF ba ell as param DE <u>REMARK</u> Signat	best spe indwidth, weter and SS Smus (72 ure/Date	ctral RF band- lyzed. FC) <u>9/17/5</u> 0

..........

. . .

Sicak times

•

Same in the

Page A-11 NO. TP-102 IN-SITU MEASUREMENTS OF CABLES AND SIGNALS TITLE FROM AREA RADIATION MONITOR HP-R-212 REV. 0 7. Using the Nicolet Model 444 FFT Analyzer (or equivalent) perform FFT analysis of signals from the following test points: SIGNAL CABINET 12 PARAMETER РНОТО # 0-100kH: 102-15 +600V *a. TB110-5 0-1k1+2 102-16 TB110-10 GND 0-100 kH2 102-17 SIG IN TB110-6 b. 0-1 KHe TB110-7 GND 102-18 TB110-8 +10V 102-19 0-100 hHz ċ. TB110-10 GND *Decouple DC voltage input to Spectrum Analyzer (50VDC Max input) If PSD plots from any one of the three signals show high or unusual amplitudes, utilize the zoom feature to provide finer resolution and obtain PSD data in the frequency band of interest. A/TEC) 9/17/80 gnature/Date 8. Inside Cabinet 12 perform usual electronic calibrations using applicable instrument shop procedures. Attach a copy of the instrument shop calibration data sheet and identify any significant adjustments in the space below: PAGE⁹ of 15



	IN-SITU MEASUREMENTS	MEASUREMENTS OF CABLES AND SIGNALS			
	TITLE FROM AREA RADIATION N	TITLE FROM AREA RADIATION MONITOR HP-R-212			
	Parate	······			
	(CABINET 12.)	SIGNAL IDENT.			
	TB110-1 (Blue)	c.s.			
	TB110-2 (Orange)	c.s.			
	TB110-3 (White)	Rem. Meter			
	TB110-4 (Yellow) IT2995C	HI N.C.			
• •	TB110-5 (RG 59/U, 72 OHM)	600V			
	TB110-6 (RG 58/U, 50 OHM)	SIG IN			
	TB110-7 (RG 58/U, 50 OHM)	Shield (for signal)	37 		
	.TB110-8 (Red)	+10V	,		
	TB110-9 (Green) IT2996C	Alert N.C.			
:	TB110-10 (B1k) (RG 59/U, 72 OHM)	GND Shield			

Signature/Date (TEC) 9/18/30

11. Using the Hewlett-Packard Model 4274 (or equivalent) Impedance Bridge measure the capacitance and impedance of the following test points:

4.4

	TITLE FROM AREA RADIATION MONITOR HP-R-212					
			· · ·			
TEST POINT		(RED) FROM*		(<i>B</i> lock) TO*		
	CABLE	WIRE COLOR/TYPE	CABLE	WIRE COLC	R/TYPE	
a.	1718701	Blue (1)	IT18701	Orange (2)		
b.	IT1870I	RG 59/U Center (5)	IT18701	RG 59/U Sh1	eld (10)	
¢	IT1870I	RG 58/U Center (6)	1718701	RG 58/U Shi	eld (7)	
d	IT1870I	Red (8)	1718701	Black (10)		
e.	IT1870I	Black (10; Field Side)	IT1870I	TB109-10 (C	abinet Side)	

*Numbers in parentheses refer to TB110 terminal numbers (field side).

Record the data required below:

	Test Point*	Cap	acitance			Impedance		WARINBLE
	Frequency	100 Hz	1 kHz	100 kHz	100 Hz	1 kHz	100 kHz	1
<u>Coil</u>	a. (1/2)	-9.84µF	-5. 11 µF	+838nF	69.12	75.4 JL	4.23.2	168
	b. (5/10)	400 nF	22 m F	-3.8µF	4.7 hJ	7.0 kil	2.08.2	Phise
	c. (6/7)	BESE (NOISE)	30 n F	-4,0µF	4.9 kA	4.9k.1	1.88N	
/	d. (8/10)	6µF	4.6 µF	186 n F	250 N	51 N	30.1	
1	e. (10/10)†	NOISE	- 76 µF	43 n F	12	4.4.2	,39 A.	
	*Numbers in tField side/	parentheses Cabinet side	refer to across c	TB110 FROM/ open link.	TO termina	il numbers	on field s	.ide. ≂) 9/14/21
	76.(5/10)	NOISE	20nF	195nF	VAR 4+14hAs	Signati	re/Date	
	with 10' RG test cable	58	PA	GE 12 of 15	VAR	5+)1 7HAS -14 10	330	;

	IN-SITU MEASUREMENTS OF CABLES AND SIGNAL				NO. TP-102		
		TITLE FROM AREA RADIATION MONITOR HP-R-212		REV.			
12,	Using t ments o	the Tektron on the five	ix Model 1 test poin	.502 (or e its given	quivalent) TDR u in Step 11. Rec	nit perform TDR ord data below:	measure-
	Te	est Point*	High R @ N ft.	Low R @ N ft.	Instrument Settings Ampl Range Mu	Strip Chart It Number	
	a.	(1/2)		·	500mp 100th 1	102-1	5
•	b	(5/10)				102-2	
2	c.	(6/7)		Constantino Antonio de la constantino Antoni	11	102-3	
1	d.	(8/10)			11	102-4	1 ga
	e	(10/10)†			<i>H</i>	102-5	
	*N(5	umbers in p ide). ield side/0	arentheses abinet sid	refer to	TB110 FROM/TO to	erminal numbers	(field
				, «			
;			-				
			:			Signature/Dat	(Tec) 9/12 e
13.	Using t ments c	he Keithle In the Test	y Model 14 Points sp	4 (or equ ecified an	ivalent DMM) per nd record value	form resistance in space provid	measure- ed.
			•	, [,] ,	;•••		
							•
				· · ·	:	,	
	÷				ti Ang ti ang		
				PAGE 13	of 15	, , ,	

	TITLE	Page A-1 IN-SITU MEASU	6 REMENTS OF CABLES AND	SIGNALS	NO. TP-102
					3EV.
			20 k.)	SCALE	· · · · · · · · · · · · · · · · · · ·
	• •		POLARITY	POLARIT	(
т е да			From = +; To = -	From = -; To	; = +
TEST POINT	FROM LINK (field side)	TO LINK (field side)	RESISTANCE	RESISTAN	ICE
a. b.	TB110-1 TB110-1 TB110-1	TB110-2 TB110-5 TB110-5	38A ~~	40.r ~	
d. e.	TB110-1 TB110-1	TB110-7 TB110-8	\sim	\sim	51 6 m
f. g. h.	TB110-1 TB110-2 TB110-2	TB110-10 TB110-5 TB110-6	~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~	~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~	¢
<u>1.</u> J.	TB110-2 TB110-2	TB110-7 TB110-8	~	\sim	
k. 1. m	TB110-2 TB110-5	TB110-10 TB110-6	\sim		
n. o.	TB110-5 TB110-5	TB110-8 TB110-10	\sim	\sim	
p. q.	TB110-6 TB110-6 TB110-6	TB110-7 TB110-8 TB110-10	9.94 hr 8.7 hr	7.4kJ 7.1kJ	
s. t. u.	TB110-7 TB110-7 TB110-7 TB110-3	TB110-8 TB110-10 TB110-10 TB110-10	7.7 kA 1 N 15.2 kA	15.2 kJ 10 7,65 k	r.
NOTE: Close all li	nks on TB110	(opened in St	ep 10) when finished	with this step),
					TEC \ 9/1.
			ST.	gnature/Date	

, 3

•		IN-SITU MEASUREMENTS OF CABLES AND SIGN	ALS TP-102
		TITLE FROM AREA RADIATION MONITOR HP-R-212	REV.
	14. Notify Shift	Supervisor/Shift Foreman of end of test on HP-R-	212.
	I hereby certify t all data has been	that this Test Procedure has been completed as wr correctly entered and filed as requested.	itten and that
- :		TEC Representative	Iner (TEC) 9/18 10/Date
		Instrumentation T.P.	[]] DEA 9/18/50
	. •		
			· . •
-			
5 - - -			· · · · · · · · · · · · · · · · · · ·
		PAGE 15 of 15	

Page A-18

JOB TICKET (WORK REQUEST) REVIEW - CLASSIFICATION - ROUTING CONTROL FORM

1	JOB TICKET NUMBER_		
1.	Does work represent a change or modification to an existing system or component? If yes, an ap-		
	proved change modification is required per AP 1021. C/M No	Yes	No
2a.	Does work requires an RWP?	Yes	No
2 b.	Is an approved procedure required to minimize personnel exposure?	Yes	No
3a.	Is work on a OC component as defined in GP 1008?	Yes	No
<u>3</u> 6.	If 3a is yes does work have an effect on Nuclear Safety? If 3b is yes, PORC reviewed Superinten- dent approved procedura must be used.	Yes	No
4.	Agreement that a PORC reviewed, Superintendent approved procedure is not required for this work because it has no effect on nuclear safety. (Applies only if 3a is Yes and 3b is No).		
ŧ	UNIT SUPERINTENDENT DATE		
5a.	Is the system on the Environmental Impact list in AP 1026?	Yes_	No
5b.	If 5a is YES, is an approved procedure required to limit environmental impact?	Yes_	No
6.	Agreement that 5b is No. (Required only if 5a is Yes).		
	UNIT SUPT ISUPY OF OPERATIONS DATE		
7.	Plant status or prerequisite conditions required for work. [Operating and/or shutdown]		
8.	OC Dept. review, if required in item No. 3.		
-	OC SUPERVISOR DATE	-	
9.	Does work require code inspector to be notified?	Yes	No
۵.	Supervisor of Majotenance approval to commence work:		
	Date 9/12/80		
۱.	Maintenance Foreman Assigned:		
2.	Code Inspector Notified. Name:	Date	
	Shift Foreman's approval to commence work:	Date 🔗	link
:			

GENERATION CORRECTIVE MAINTENANCE SYSTEM CM STATUS ACTIVITY FORM Page A-19



(

Page A-20

MTX-501

Toler.

±.5V

±.1V

±3.0V

Toler.

Area Monitor •

4P-R-212

DETECTER

RATEMETER

Model____ Serial___

Model_____ Serial___

-. 3

			-	
FCK Posit.	Desired Mr./Hr.	As Found	As Left	⁻ Toler.
Closed		••		
Iter.		N/	4	
Open				

Check Source Rdg. ____Mr/Hr

Fail Safe_____Volts

Section

Ratemeter

Mr/Hr	Desired Mv. Out	As Found	As Left	Toler.		Pwr. Supply	Found	As Left	To
104	1.000	+.9892		.150	Ţ	-6.8V	-6.6		+.
10 ³	. 8007	7. 7892		.151		10.00	1.1.08		÷.
10 ²	.600V	1.5958		.151		22.OV	+19.37		<u>±3</u>
10	. 400V	· 4000		.154		•	•*		-
1	.200V	+. 2081		.150		Alarm Set. Pt.	As Found	As Left	То
.1	.000V	7.0174		.151					
EQUIP.	Tuke Sia	SER.NO.	530314 1200-6	LAS	тс	AL. 7-2 AL. 8-1	5-80 9-80	DUE <u>/-25</u> DUE <u>2-1</u>	T-81 1-91
EQUIP.	QUIPSER.NO			LAST CAL				DUE	
PERFORMED	BY Til	. Helber	DATE 9	-15-20	AP	PROVED BY		DA	TE



STRIP CHART 102-5 TB110: 10 to 10

Signal - GND

