

# **Blind QC trend presentation**

October 2024

Julia Elizabeth Hutto Strock





#### DISCLAIMER

This information was prepared as an account of work sponsored by an agency of the U.S. Government. Neither the U.S. Government nor any agency thereof, nor any of their employees, makes any warranty, expressed or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness, of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. References herein to any specific commercial product, process, or service by trade name, trade mark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the U.S. Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the U.S. Government or any agency thereof.

## **Blind QC trend presentation**

Julia Elizabeth Hutto Strock

October 2024

Idaho National Laboratory Idaho Falls, Idaho 83415

http://www.inl.gov

Prepared for the U.S. Department of Energy Under DOE Idaho Operations Office Contract DE-AC07-05ID14517

October 23, 2024 **Julia Strock** Indirect Radiobioassay Technical Lead **Blind QC Trends** 

## **Value of Blind Performance Testing**

- Assess the analytical performance on a regular basis
- Evaluate measurement capabilities
- Ensure quality assurance objectives are met
- Identify Trends
- Take action for improvement
- Meet DOELAP requirements

## **INL's Blind Performance Testing Program**

- Double blind authentic urine and fecal samples are sent to the analytical laboratory quarterly. DOE-RESL site-specific performance evaluation program (SSPEP) spikes the samples per INL's requested parameters.
- Typical Schedule

| <u>Urine</u>       |                        | <u>Fecal</u>       |                        |
|--------------------|------------------------|--------------------|------------------------|
| Quarterly          | Annually               | Quarterly          | Annually               |
| Am-241             | Np-237                 | Am-241             | Sr-90                  |
| Pu-238, Pu-239/240 | Th-228, Th-230, Th-232 | Pu-238, Pu-239/240 | Np-237                 |
| H-3                | U-233/234, U-238       |                    | Th-228, Th-230, Th-232 |
| Total U            | Gamma(Co-60, Cs-137)   |                    | U-233/234, U-238       |
|                    | Sr-90                  |                    | Gamma(Co-60, Cs-137)   |

# **Performance Evaluation Acceptance Criteria**

| Blind Performance Evaluation<br>Type | Individual Sample Acceptance Criteria                  |
|--------------------------------------|--|
| Accuracy and Agreement               | Bias% = +/-30% and < 3 sigma (99% confidence interval) |
| Sensitivity                          | Agreement < 3  |
| False Positive                       | Result - (3*CSU) < 0                                   |
| False Negative                       | Result +/- (3*CSU) > 0                                 |

# **Out of Control Signals for Trending Analysis**


| Blind Performance<br>Evaluation Type      | Trending Analysis Out of<br>Control Signals  | Signal Indicated  |
|---|--|---|
| Accuracy                                  | <ol> <li>3 sequential points outside +/- 15% average bias.</li> <li>9 sequential points on one side of the centerline without crossing.</li> <li>6 points in a row steadily increasing or decreasing.</li> </ol> | A medium tendency for out of control conditions.     A prolonged bias exists.     A trend exists. |
| Agreement (for accuracy evaluations only) | 3 sequential points above 2 sigma.   | A medium tendency for out of control conditions.  |
| Precision (for accuracy evaluations only) | 3 sequential results > 0.40.   | A medium tendency for out of control conditions.  |

### References for Criteria:


- ANSI N13.30-2011(R2017
- RESL-PLN-1
- MAPEP-HB-1
- Nelson rules of process control

## **Americium-241 in Urine Bias Trends**

· Average Bias has 9 sequential points on the negative side of the centerline without crossing



Individual Bias results



$$B_r = \sum_{i=1}^n \frac{B_{ri}}{n}$$

Where:

 $B_{ri}$  = individual relative bias

n = the number of measurements (sample size is at least five)

$$Bias = \frac{C_m}{C_l} - 1$$

Where:

 $C_m$  = the measured activity of the control

 $C_t$  = the true (known) activity of the control.

# **Am-241 in Urine Agreement Trend**


### Agreement

$$|A| = \frac{(C_m - R_r) - 1}{CSU_r}$$

#### Where:

|A| = agreement (absolute value)  $C_m =$  measured activity of the control  $R_r =$  reference value ratio of measured activity and known activity  $CSU_r =$  CSU of ratio

- < 3 sigma=99% confidence interval; <2 sigma=95% confidence interval
- Out of control signal: 3 sequential points above 2 sigma agreement




warning limit = 2 sigma control limit = 3 sigma

### Possible Causes?

- Non-homogeneous matrix
- Age of samples

## Am-241, Pu-238, Pu-239/240, and Sr-90 in fecal matrix

• Am-241, Pu-238, Pu-239/240, and Sr-90 in fecal matrix have 9 sequential points on the negative side of the centerline without crossing.




#### Possible Causes?

This trend is theorized to be due to incomplete dissolution of solids which could be related to the age of the spiked samples. Our stock of blind PE fecal samples has been in the freezer for over 3 years. Remaining stock of blind PE fecal samples have been analyzed, so future blind samples will be freshly spiked to eliminate the age of the samples as a contributing factor to out-of-control signals.

# **Total Uranium in Urine Agreement**


Total Uranium has 3 sequential points above 2 sigma agreement.






### Possible Causes?

- Bias Fluctuation
- Spike Levels>20\*RDL
- Age of samples



### **Other Trends**

- Cs-137 in urine has exhibited positive bias trends. Recent blind testing included false positive and sensitivity evaluations that all passed acceptance criteria.
- Co-60 in urine has exhibited positive bias trends. Recent blind testing included false positive, false
  negative, and sensitivity evaluations. One false negative test out of the last two failed acceptance
  criteria. However, the result was greater than our critical level. If it were a worker's sample, followup recounts would have been requested.
- H-3 in urine showed a positive trend of 9 sequential results above 0% bias until the most recent results.



## **Discussion**





Battelle Energy Alliance manages INL for the U.S. Department of Energy's Office of Nuclear Energy. INL is the nation's center for nuclear energy research and development, and also performs research in each of DOE's strategic goal areas: energy, national security, science and the environment.